

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 3-16-2026

Instrument #: 100358

Instrument Location: TOK POST,

Supervisor Name and Agency: SGT. SEAS HOSSLER, AST

Describe the Status Change or Issue with the Instrument:

UNIT WAS GIVING FREQUENT RFI, INTERNAL  
STANDARD ERRORS.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time:

Out of Service Date/Time: 3-16-2026, 1645 HRS

For Use by SCDL

Additional Notes

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov).

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor