

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument  
State of Alaska

Serial #: 100420

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/26/2024

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 28.34: 0.076  
LOT #: AG310901  
EXPIRATION: 04/19/2025  
TANK PRESSURE: 939 psi

BLANK TEST	0.000	09:23
INTERNAL STANDARD	VERIFIED	09:23
EXTERNAL STANDARD	0.076	09:24
BLANK TEST	0.000	09:24
EXTERNAL STANDARD	0.075	09:25
BLANK TEST	0.000	09:26
EXTERNAL STANDARD	0.075	09:26
BLANK TEST	0.000	09:27
EXTERNAL STANDARD	0.075	09:27

Filter Wheel Error

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.8°C	PASSED
Breath Tube = 48.1°C	PASSED

### PUMP INFO

Flow Rate = 4.678 L/M	PASSED
-----------------------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

File only dm 08/21/24