

Alaska Scientific Crime Detection Laboratory

DataMaster Calibration Form

Issued: 4/12/2018
Effective: 4/13/2018

Version: DMCAF 2018 R0
Status: Active

Instrument # 100374

Analyst Brandi Barnett

Date: 10/29/19

Reason for Calibration

Interference Detected on Tanks in field

1. Wet Bath Simulator Solution

- a. Level 0.100
- b. Manufacturer Guth
- c. Lot 10370
- d. Expiration Date 12/5/20

2. Wet Bath Simulator

- a. Serial Number ^{BMB} SD198
- b. Date Calibrated 3/6/19

3. Attach Calibration Factors

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100374

Date: 10/29/2019
Time: 12:34:32

OPERATOR NAME:
BRANDI BARNETT
CERTIFICATE #: 7038

Ca = 0.1000
CAL = 0.949489 0.800 <= CAL < 1.200
b1 = 0.0004 0.0000 <= b1 < 0.0040
b2 = 0.0054 0.0010 <= b2 < 0.0100
b3 = 0.0004 0.0000 <= b3 < 0.0040
Xq = 0.0944 0.0500 <= Xq < 0.2000
a21 = 1.211616 1.050 <= a21 < 1.300
a31 = 0.395006 0.300 <= a31 < 0.800