	State of Alaska Scientific Crime Detection Laboratory		
	<b>DataMaster Acceptance Form - DMAF2011R0</b>		
	Issuing Authority - Nita Bolz, Scientific Director		
Approval date: 03/01/11	Effective date: 03/01/11	Archive date: Active	

**DataMaster DMT Acceptance Form**

Instrument # 100667

Date: 10/12/11

Analyst: CTF

1. Check in instrument in to JusticeTrax Breath Alcohol Database
2. Visual Inspection
3. Update software
4. Print calibration factors (attached)
5. NPAS calibration paperwork (attached)
6. Barcode Instrument
7. State of Alaska Property ID# 12-F0358

**B-100667** 100667




QSHS03C20UNPNH B-100667  
Breath Program

## CALIBRATION FACTORS

-----  
Alaska Department of Public Safety  
DATAMASTER dmt: 100667  
-----

Date: 09/21/2011  
Time: 09:08:19

Ca	=	0.1510			
CAL	=	0.991475	0.800	<=	CAL < 1.200
b1	=	0.0001	0.0000	<=	b1 < 0.0040
b2	=	0.0036	0.0010	<=	b2 < 0.0100
b3	=	0.0000	0.0000	<=	b3 < 0.0040
Xq	=	0.1075	0.0500	<=	Xq < 0.2000
a21	=	1.219071	1.050	<=	a21 < 1.300
a31	=	0.470220	0.300	<=	a31 < 0.800



**National Patent  
Analytical Systems, Inc.**

*Certificate of Calibration*

100667

Instrument Serial Number

11030

Standard Reference Number

The above instrument has been calibrated and is hereby certified for use in determining alcohol breath concentration when used in accordance with the operating guidelines by a properly trained operator.

9/21/11

Date

Charles E. Townsend

Technician

Final page for report

\

\\ANCSCLSAN01\JTRAX\CASEDOCS\B-100667\B-100667 ECF-20140

7/10/2014