

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 6/11/2014  
Effective: 6/11/2014

Version: CIISF 2014 R0  
Status: Active

Date: 09/11/15 Instrument Serial Number: 100365  
~~100688~~ <sup>9/21/15</sup> <sub>COB 9/11/15 08</sub>

Instrument Location: 2580 Fourth Ave

Supervisor Name and Agency: James Re Kotzebue Police Dept

Describe the Status Change or Issue with the Instrument:

Internal standard errors, also each time  
I push the DM7 Icon it causes an error  
and reboots. unable to take machine out  
of service

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: 9-11-15 @ 1120

For Use by SCDL

Additional Notes

Instrument removed from service in Kotzebue was  
100365. I corrected the serial number.  
COB 9/21/15

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740