

VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument
State of Alaska*

Serial #: 100381

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 03/17/2021

External Standard Test Values

Diagnostic Check

VERSIONS

DMT: 3.02

PIC: 3.02

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 49.0°C

PASSED

Breath Tube = 47.8°C

PASSED

PUMP INFO

Flow Rate = 4.670 L/M

PASSED

RFI DETECTED

file only JMC 04/14/21