

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100385

Received From: AST/Fairbanks

Date: 12/14/22
COB

Analyst: C. O'Bryant

Reason for instrument return:

Removed from service for "unable to
load ticket db" pop up box
Screen isn't recognizing touch.

Change instrument and case location in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax

Upload Unable to upload due to ticket db msg.*

Update agency information and polling status in DMHost Calling Program

Check for VOCs

From Intoximeters

Intoximeters Repair Paperwork (attached)

** successfully uploaded during
repair + checked for VOCs.
COB 12/14/22*