

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 12/15/2017  
Effective: 12/15/2017

Version: CIISF 2017 R0  
Status: Active

Date: 10/15/19 Instrument Serial Number: 100381

Instrument Location: Soldotna AST

Supervisor Name and Agency: Ronny Simmons / AST

Describe the Status Change or Issue with the Instrument:

Put this Detester into service

Time was off 1 hr at set up  
it was really 0952, it has now been changed

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 10/15/19 / 1057

Out of Service Date/Time: \_\_\_\_\_

For Use by SCDL

Additional Notes

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740