

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100365

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 03/06/2025

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 28.98: 0.077
LOT #: AG310901
EXPIRATION: 04/19/2025
TANK PRESSURE: 568 psi

BLANK TEST	0.000	23:07
INTERNAL STANDARD	VERIFIED	23:08
EXTERNAL STANDARD	0.077	23:08
BLANK TEST	0.000	23:09
EXTERNAL STANDARD	0.076	23:09
BLANK TEST	0.000	23:10
EXTERNAL STANDARD	0.076	23:10
BLANK TEST	0.000	23:11
EXTERNAL STANDARD	0.076	23:11
BLANK TEST	0.000	23:12
EXTERNAL STANDARD	0.076	23:12
BLANK TEST	0.000	23:13

Average = 0.0762
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C	PASSED
Breath Tube = 45.2°C	PASSED

PUMP INFO

Flow Rate = 4.533 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

File Only DM 06/26/25