

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100391

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 10/15/2019

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.18: 0.078
LOT #: AG826101
EXPIRATION: 09/18/2020
TANK PRESSURE: 736 psi

BLANK TEST	0.000	09:06
INTERNAL STANDARD	VERIFIED	09:06
EXTERNAL STANDARD	0.077	09:06
BLANK TEST	0.000	09:07
EXTERNAL STANDARD	0.077	09:07
BLANK TEST	0.000	09:08
EXTERNAL STANDARD	0.077	09:08
BLANK TEST	0.000	09:09
EXTERNAL STANDARD	0.077	09:09
BLANK TEST	0.000	09:10
EXTERNAL STANDARD	0.077	09:11
BLANK TEST	0.000	09:11

Average = 0.0770
Std Dev = 0.0000

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.01
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C PASSED
Breath Tube = 48.0°C PASSED

PUMP INFO

Flow Rate = 4.480 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD

PASSED

File only: Instrument being removed
from service. BMB 10/23/19