

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: March 22, 2019 Instrument Serial Number: 100347

Instrument Location: Craig Police Department, 506 Second Street

Supervisor Name and Agency: Roberto Medina / Craig Police Department

Describe the Status Change or Issue with the Instrument:

Instrument kept giving error messages. Chief Ely, with phone support attempt to correct the problem, but wasn't able to. Instrument taken off line, out of service and new instrument shipped

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: March 22, 2019 at 12:42 PM

For Use by SCDL	Additional Notes
	<u>Repeated gas flow errors. DM 04/04/19</u>

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740