

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Issued: 04/03/2018  
Effective: 04/04/2018

Version: DMEF2018 R0  
Status: Active

Instrument # 100362

Date: 12/3/19

Analyst: Brandi Barnett

1. Reason for instrument evaluation

Preventative Maintenance

2. Visual Inspection

Clean

Replace missing screws

Cermetex Box is secure

Ensure modem cable present

Gas Valve upright or secured to wall

All tubing and connections are secure

3. O-ring inspection and ensure 2 O-rings in replacement bag

4. Troubleshooting/Repair

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

5. VOC (attached)

6. Internal Standard value  $\leq 3.00\%$  Yes  No

7. Verify software version

Return to Factory for Repair

Ready for Calibration

Ready for Certification

Proceed to Close Out

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100362

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/03/2019

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.43: 0.079  
LOT #: AG826101  
EXPIRATION: 09/18/2020  
TANK PRESSURE: 613 psi

BLANK TEST	0.000	12:16
INTERNAL STANDARD	VERIFIED	12:16
Xq = 0.072 (0.60%)		
EXTERNAL STANDARD	0.078	12:16
X[1] = 0.0780 (-0.0040) (-0.0005)		
BLANK TEST	0.000	12:17
EXTERNAL STANDARD	0.079	12:17
X[1] = 0.0785 (-0.0035) (-0.0005)		
BLANK TEST	0.000	12:18
EXTERNAL STANDARD	0.078	12:18
X[1] = 0.0780 (-0.0039) (-0.0005)		
BLANK TEST	0.000	12:19
EXTERNAL STANDARD	0.078	12:20
X[1] = 0.0776 (-0.0039) (-0.0010)		
BLANK TEST	0.000	12:20
EXTERNAL STANDARD	0.078	12:21
X[1] = 0.0781 (-0.0036) (0.0001)		
BLANK TEST	0.000	12:22

Average = 0.0782  
Std Dev = 0.0004

## Diagnostic Check

VERSIONS  
DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.6°C	PASSED
(44.0°C - 52.0°C)	
Breath Tube = 45.1°C	PASSED
(38.0°C - 50.0°C)	

### SETTINGS

Lamp Voltage = 1.77 V	PASSED
(1.10 V - 2.56 V)	
Cooler Voltage = 1.87 V	PASSED
(0.97 V - 2.25 V)	
Bias Voltage = 79 V	PASSED
(47 V - 111 V)	
Chopper Freq = 523 Hz	PASSED
(475 Hz - 575 Hz)	
Barometer = 29.4 in	

### PUMP INFO

Flow Rate = 5.208 L/M	PASSED
(3.500 L/M - 6.500 L/M)	

### DETECTOR INFO

PUMP ON	PASSED
(0.002 V <= 0.010 V)	
PUMP OFF	PASSED
(0.002 V <= 0.010 V)	

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

### INTERNAL STANDARD

Xq = 0.072	0.71%	(0.00% - 4.00%)	PASSED
------------	-------	-----------------	--------