

	State of Alaska Scientific Crime Detection Laboratory		
	DataMaster Calibration Form – DMCAF2011R1		
	Issuing Authority – Nita Bolz, Scientific Director		
	Approval date: 03/10/11	Effective date: 03/10/11	Archive date: Active

DataMaster DMT Calibration Form

Instrument # 100363

Date: 4/26/11

1. Wet Bath Simulator Solution

- a. Level 0.100
- b. Manufacturer Buth
- c. Lot 10260
- d. Expiration Date 10/25/12

2. Wet Bath Simulator

- a. Serial Number DR4161
- b. Date Calibrated 8/23/10

3. Attach Calibration Factors

Breath Alcohol Analyst C. Bryant

COB/2

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100363

Date: 04/26/2011
Time: 08:24:29

SUPERVISOR NAME:
COLLEEN S OBRYANT
CERTIFICATE #: 6237

LOT #: GUTH100
BOTTLE #: 10260
EXPIRATION: 10/25/2012

Ca = 0.1000
CAL = 0.982000 0.800 <= CAL < 1.200
b1 = 0.0002 0.0000 <= b1 < 0.0040
b2 = 0.0049 0.0010 <= b2 < 0.0100
b3 = 0.0003 0.0000 <= b3 < 0.0040
Xq = 0.0808 0.0500 <= Xq < 0.2000
a21 = 1.164486 1.050 <= a21 < 1.300
a31 = 0.451262 0.300 <= a31 < 0.800

COB/A