

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: January 6, 2018 Instrument Serial Number: 100363

Instrument Location: Nome, AK

Supervisor Name and Agency: Nicholas Harvey - Nome Police Department

Describe the Status Change or Issue with the Instrument:

Filter wheel error that was not corrected by troubleshooting steps. New Datamaster DMT sent and set up, with the malfunctioning unit being returned to the Alaska Scientific Crime Detection Laboratory.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: December 24, 2017 / 0709 hours

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740