

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 1/9/2025

Instrument #: 100349

Instrument Location: Metlakatla Police Department

Supervisor Name and Agency: Justin Janes Metlakatla Police Department

Describe the Status Change or Issue with the Instrument:

I was contacted by the Alaska Crime lab notifying me that there had been instances of Internal Standard errors occurring with the DataMaster and were sending my agency a replacement. I took the old instrument offline and placed the new one online.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 01/09/2025 0004 Hours

Out of Service Date/Time: [Click or tap here to enter text.](#)

For Use by SCDL

Additional Notes

100398 was taken offline as a result of the 2024.10.25 – 2024.12.12 In Field Instrument Review due to internal standard errors. 100349 was sent to replace this instrument. MNC 1/22/25

Both change in instrument status forms were sent to the lab as one combined document, the information was split into separate documents and the instrument numbers were corrected. The original is attached to page 2. MNC 1/22/25

Email completed form to dps.scdl.toxicology@alaska.gov.

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 1/9/2025

Instrument #: Alaska 3-22860

Instrument Location: Metlakatla Police Department

Supervisor Name and Agency: Justin Janes Metlakatla Police Department

Describe the Status Change or Issue with the Instrument:

I was contacted by the Alaska Crimelab notifying me that there had been instances of Internal Standard errors occurring with the DataMaster and were sending my agency a replacement. I took the old instrument offline and placed the new one online.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 01/09/2025 0004 Hours

Out of Service Date/Time: 01/08/2025 2322 Hours

For Use by SCDL

Additional Notes

Click or tap here to enter text.

Email completed form to dps.scdl.toxicology@alaska.gov.

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor