

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 1/13/2017
Effective: 1/17/2017

Version: DMCIF 2017 R0
Status: Active

Instrument S/N 100397

Received From APD

Date 9/12/17

Analyst Brandi Barnett

1. Reason for instrument return:

Instrument being used by APD temporarily for Fair.

2. Change instrument and case location in JusticeTrax Breath Alcohol Database
3. Ensure change in instrument status form is in JusticeTrax
4. Upload
5. Update agency information and polling status in DM Host Calling Program
6. Check for VOCs

From Intoximeters

7. Intoximeters Repair Paperwork (attached)