

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 12/15/2017  
Effective: 12/15/2017

Version: CIISF 2017 R0  
Status: Active

**JUL 10 2018**

Date: 07/10/2018 Instrument Serial Number: 100666

Instrument Location: Bethel PD (157 Salmonberry St.), in Patrol room.

Supervisor Name and Agency: Justin L. Ulak - Bethel PD

Describe the Status Change or Issue with the Instrument:

Changing Instrument due to Internal Standard Errors received.

---

---

---

---

---

---

---

---

---

---

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: 07/10/2018 at 12:45pm

-----  
**For Use by SCDL**

**Additional Notes**

---

---

---

---

---

---

---

---

---

---

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740