

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100387

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/16/2019

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 29.71: 0.079

LOT #: AG923401

EXPIRATION: ~~07/22/2021~~ 08/22/2021 DM 12/26/19

TANK PRESSURE: 1165 psi

BLANK TEST	0.000	19:29
INTERNAL STANDARD	VERIFIED	19:29
EXTERNAL STANDARD	0.080	19:29
BLANK TEST	0.000	19:30
EXTERNAL STANDARD	0.080	19:30
BLANK TEST	0.000	19:31
EXTERNAL STANDARD	0.080	19:31
BLANK TEST	0.000	19:32
EXTERNAL STANDARD	0.080	19:33
BLANK TEST	0.000	19:33
EXTERNAL STANDARD	0.080	19:34
BLANK TEST	0.000	19:35

Average = 0.0800

Std Dev = 0.0000

## Diagnostic Check

### VERSIONS

DMT: 3.02

PIC: 3.02

Modem: 2.6

Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.8°C PASSED

Breath Tube = 47.7°C PASSED

### PUMP INFO

Flow Rate = 4.578 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED

PUMP OFF PASSED

### FILTER INFO

Filter 1 PASSED

Filter 2 PASSED

Filter 3 PASSED

INTERNAL STANDARD

PASSED

I, Brandi M. Barnett, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.


(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.


(4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.



(5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Brandi M. Barnett  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 10 day of Jan, 2020

  
Notary Public  
My Commission Expires With Office



DM 12/26/19