

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100690

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 06/04/2019

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 30.21: 0.081
LOT #: AG826101
EXPIRATION: 09/18/2020
TANK PRESSURE: 996 psi

| | | |
|-------------------|----------|-------|
| BLANK TEST | 0.000 | 13:37 |
| INTERNAL STANDARD | VERIFIED | 13:37 |
| EXTERNAL STANDARD | 0.077 | 13:37 |
| BLANK TEST | 0.000 | 13:38 |
| EXTERNAL STANDARD | 0.078 | 13:38 |
| BLANK TEST | 0.000 | 13:39 |
| EXTERNAL STANDARD | 0.079 | 13:39 |
| BLANK TEST | 0.000 | 13:40 |
| EXTERNAL STANDARD | 0.078 | 13:41 |
| BLANK TEST | 0.000 | 13:41 |
| EXTERNAL STANDARD | 0.079 | 13:42 |
| BLANK TEST | 0.000 | 13:43 |

Average = 0.0782
Std Dev = 0.0008

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
Breath Tube = 48.1°C PASSED

PUMP INFO

Flow Rate = 4.321 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED

I, Charles R. Foster, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.


(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.


(5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.


(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

 7/25/19

Charles R. Foster
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 25 day of 07, 2019


Notary Public
My Commission Expires With Office



COR 7/15/19