

In Field Instrument Review
Dates: 07/18/2018 – 09/05/2018
Reviewed by Colleen O’Bryant

100255 (Clear AFB) – 1/2 – RFI – No action taken
100349 (Ketchikan PD) – 1/15 – Pump Error – No action taken
100351 (Fairbanks Airport PD) – 1/10 – RFI – No action taken
100352 (Anchorage Airport PD) 2/5 – Ambient Fail, Detector Overflow – both messages were conducted during “test,test”, no action taken
100355 (Seward PD) – 1/16 – RFI – No action taken
100360 (AST/Soldotna) – 1/17 – Internal Standard Error – message followed by valid test, will monitor instrument.
100374 (Anchorage PD) 14/126 – RFI, Pump Error, Interference, Tank Interference, Blank Error, Ambient Fail –Tank change performed on 09/14/2018 - No action taken
100378 (AST/Fairbanks) – 4/30 – Pump Error, Ambient Fail – No action taken
100379 (Wasilla PD) – 7/45 – RFI, Pump Error, Interference – No action taken
100380 (Kenai PD) - 2/21 – Blank Error, Ambient Fail – No action taken
100382 (UAF PD) – 1/5 – Pump Error – No action taken
100385 (AST/Ketchikan) – 2/3 – RFI, Ambient Fail – No action taken
100387 (Petersburg PD) – 1/6 – Blank Error – No action taken
100395 (Hoonah PD) –7/15 – Blank Error, Gas Flow Error, Ambient Fail, Static Analysis Error - Removed from service
100402 (Anchorage PD) – 12/136 – RFI, Pump Error, Interference, Blank Error, Standard Out of Range, Ambient Fail - Removed from service
100404 (St. Paul PD) – 3/3 – RFI – Contact agency regarding location of potential RFI source.
100406 (Juneau PD) – 3/25 – RFI, Pump Error, Interference – No action taken
100416 (Fairbanks PD) - 1/46 – Ambient Fail – No action taken
100422 (JBER) – 1/7 – RFI – No action taken
100423 (Ft. Wainwright MP) – 2/16 – RFI, Interference – No action taken
100669 (AST/Tok) – 1/9 – RFI – No action taken
100674 (Juneau PD) – 3/13 – RFI, Interference, Ambient Fail – No action taken
100688 (Palmer PD) – 6/39 – RFI – No action taken
100689 (Kotzebue PD) – 1/10 – Tank Interference - No action taken
100690 (Unalaska PD) – 5/20 – RFI, Tank Interference, Blank Error – No action taken
100692 (AST/Aniak) – 2/8 – Tank Interference – Both instances prior to high breath samples. Monitor instrument, current tank is 1000+PSI.
100695 (JBER) – 3/14 – RFI, Interference, Blank Error – No action taken
100696 (Soldotna PD) – 1/10 – Ambient Fail – No action taken

Not included in this review: 100377 (AST/St. Mary’s), 100388 (Girdwood), 100397 (NSBPD), 100424 (Eielson), 100681 (Whittier PD)

