

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100361

Received From: NOME PD

Date: 04/10/23

Analyst: Tatiana Malnoske

Reason for instrument return:

low external standard values on 02/21/23 VOC
and multiple fatal error and interference detected
status messages.

Change instrument and case location in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax

Upload

Update agency information and polling status in DMHost Calling Program

Check for VOCs

From Intoximeters

Intoximeters Repair Paperwork (attached)