

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100678

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 05/30/2021

## External Standard Test Values

## Diagnostic Check

VERSIONS  
DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C PASSED  
Breath Tube = 48.1°C PASSED

### PUMP INFO

Flow Rate = 4.821 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED

Detector Overflow

File only DM 06/11/21