

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100350

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 10/28/2019

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 30.10: 0.080
LOT #: AG826101
EXPIRATION: 09/18/2020
TANK PRESSURE: 958 psi

| | | |
|-------------------|----------|-------|
| BLANK TEST | 0.000 | 19:37 |
| INTERNAL STANDARD | VERIFIED | 19:37 |
| EXTERNAL STANDARD | 0.079 | 19:37 |
| BLANK TEST | 0.000 | 19:38 |
| EXTERNAL STANDARD | 0.079 | 19:38 |
| BLANK TEST | 0.000 | 19:39 |
| EXTERNAL STANDARD | 0.078 | 19:40 |
| BLANK TEST | 0.000 | 19:40 |
| EXTERNAL STANDARD | 0.078 | 19:41 |
| BLANK TEST | 0.000 | 19:42 |
| EXTERNAL STANDARD | 0.078 | 19:42 |
| BLANK TEST | 0.000 | 19:43 |

Average = 0.0784
Std Dev = 0.0005

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.1°C PASSED
Breath Tube = 46.2°C PASSED

PUMP INFO

Flow Rate = 4.380 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED

I, Brandi M. Barnett, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.



Brandi M. Barnett
Scientific Director
State Breath Alcohol Program

12/13/19

Subscribed and sworn before me this 13th day of Dec, 20 19



Notary Public
My Commission Expires With Office



JW 12/9/19