

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: Click or tap to enter a date. 10/07/25

Instrument #: Click or tap here to enter text. 100354

Instrument Location: Click or tap here to enter text. SCDL

Supervisor Name and Agency: Click or tap here to enter text. Derek Walton - SCDL

Describe the Status Change or Issue with the Instrument:

Click or tap here to enter text.

Repeated Screen freezing during Breath Test Supervisor school. Filter wheel solenoid wires are badly flaking. Instrument calibration from June 2025 is voided.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: Click or tap here to enter text. NA

Out of Service Date/Time: Click or tap here to enter text. NA

For Use by SCDL

Additional Notes

Click or tap here to enter text.

Email completed form to dps.scdl.toxicology@alaska.gov.