

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100374

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 10/10/2019

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.50: 0.079  
LOT #: AG826101  
EXPIRATION: 09/18/2020  
TANK PRESSURE: 151 psi

BLANK TEST	0.000	20:40
INTERNAL STANDARD	VERIFIED	20:40
EXTERNAL STANDARD	INTERFERENCE	20:40

Interference Detected

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber =	48.8°C	PASSED
Breath Tube =	43.7°C	PASSED

### PUMP INFO

Flow Rate =	4.312 L/M	PASSED
-------------	-----------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

### INTERNAL STANDARD

PASSED