

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 6/23/2025

Instrument #: 100344

Instrument Location: Juneau PD

Supervisor Name and Agency: Derek Walton – SCDL on the phone with Sterling Salisbury – Juneau PD

Describe the Status Change or Issue with the Instrument:

While reviewing VOCs, I noticed that the lot number for the dry gas tank was entered incorrectly when it was installed on 04/24/2025. Ofc. Salisbury agreed to perform a tank change to correct this. The tank change was completed on 06/19/2025 and followed by a successful VOC. The lot number of the tank identified on the VOCs on 04/24/2025 and 05/30/2025 was hand corrected. Subject tests throughout this timeframe will continue to reflect the incorrect lot number; however, the dry gas tank lot is valid and approved for evidential use.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA - Tank changed 06/19/2025

Out of Service Date/Time: NA

-----  
**For Use by SCDL**

**Additional Notes**

Click or tap here to enter text.

**Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov).**

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor