

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument  
State of Alaska

Serial #: 100684

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/14/2013

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.64: 0.079  
LOT #: 02012080A1  
EXPIRATION: 03/01/2014

BLANK TEST	0.000	21:33
INTERNAL STANDARD	VERIFIED	21:33
EXTERNAL STANDARD	0.078	21:33
BLANK TEST	0.000	21:34
EXTERNAL STANDARD	0.078	21:34
BLANK TEST	0.000	21:35
EXTERNAL STANDARD	0.078	21:35
BLANK TEST	0.000	21:36
EXTERNAL STANDARD	0.078	21:36
BLANK TEST	0.000	21:37
EXTERNAL STANDARD	0.078	21:37
BLANK TEST	0.000	21:38

Average = 0.0780  
Std Dev = 0.0000

## Diagnostic Check

VERSIONS  
DMT: 2.00  
PIC: 3.03  
Modem: 2.3  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C PASSED  
Breath Tube = 44.8°C PASSED

### PUMP INFO

Flow Rate = 4.047 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED

### FILTER INFO


Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

### INTERNAL STANDARD

PASSED


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

12/20/13

Subscribed and sworn before me this 20 day of Dec, 20 13

  
Nikki Roth, Notary Public  
My Commission Expires With Office



CFF  
12/20/13