

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 20 May 2025

Instrument #: 100255

Instrument Location: Bldg. 37 Clear SFS AK 99704

Supervisor Name and Agency: TSgt William C. Bailey 268 SFS

Describe the Status Change or Issue with the Instrument:

Received Unit owned DataMaster

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 20 May 2025 0815

Instrument did not go into service, no passing VOC
CKG 5/20/2025

Out of Service Date/Time:

For Use by SCDL

Additional Notes

Spoke to William Bailey over the phone. The DataMaster owned by Clear AFB (100255) was returned to the agency. However, 100255 was displaying detector overflow status messages and was unable to be placed into service (no completed VOC). CKG 5/20/2025

Email completed form to dps.scdl.toxicology@alaska.gov.

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor