

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100347

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 02/08/2023

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 27.39: 0.073  
LOT #: AG113402  
EXPIRATION: 05/14/2023  
TANK PRESSURE: 480 psi

BLANK TEST	0.000	09:30
INTERNAL STANDARD	VERIFIED	09:30
EXTERNAL STANDARD	INTERFERENCE	09:31

Interference Detected

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 47.8°C	PASSED

### PUMP INFO

Flow Rate = 4.192 L/M	PASSED
-----------------------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

### INTERNAL STANDARD

PASSED

File Only DM 03/24/23