

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100699

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 05/09/2019

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.69: 0.079
LOT #: AG735001
EXPIRATION: 12/16/2019
TANK PRESSURE: 986 psi

BLANK TEST	0.000	15:37
INTERNAL STANDARD	VERIFIED	15:37
EXTERNAL STANDARD	0.074	15:37

Blank Error

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C	PASSED
Breath Tube = 39.0°C	PASSED

PUMP INFO

Flow Rate = 4.270 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

PASSED