

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 11/20/2022

Instrument #: 100692

Instrument Location: 6040 De Havilland Ave. Anchorage, AK 99502

Supervisor Name and Agency: Thomas Hoffman / Anchorage Airport Police and Fire

Describe the Status Change or Issue with the Instrument:

- 1) Rebooted the DMT. Unplugged power cable, waited 15 seconds, plugged it back in.
- 2) Plugged in wired keyboard (system driver window popped up, I clicked the "X").
- 3) Keyboard worked after this.
- 4) Successfully passed Verification of Calibration (VOC) and successfully printed.
- 5) Ran a non drinking subject test afterwards, successfully completed and printed.
- 6) Printed off an additional test strip via "copy" button. Successful.
- 7) Returned to Service after consultation with Crime Lab

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 11/20/2022

Out of Service Date/Time: Click or tap here to enter text.

For Use by SCDL

Additional Notes

Click or tap here to enter text.

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor