



State of Alaska Scientific Crime Detection Laboratory

DataMaster Calibration Form - DMCAF2011R1

Issuing Authority - Nita Boiz, Scientific Director

Approval date: 03/10/11

Effective date: 03/10/11

Archive date: Active

DataMaster DMT Calibration Form

Instrument # 100-412

Date: 6/1/11

1. Wet Bath Simulator Solution

- a. Level 0.100
- b. Manufacturer Repro
- c. Lot 10003
- d. Expiration Date 7/8/12

2. Wet Bath Simulator

- a. Serial Number DR-1161
- b. Date Calibrated 8/22/10

3. Attach Calibration Factors

Breath Alcohol Analyst

C. Bryant

CALIBRATION FACTORS

Alaska Department of Public Safety  
DATAMASTER dat: 100412

Date: 06/01/2011  
Time: 09:28:11

SUPERVISOR NAME:  
COLLEEN S O'BRYANT  
CERTIFICATE #: 6237

LOT #: REPC0100  
BOTTLE #: 10003  
EXPIRATION: 09/08/2012

Ca	=	0.1000		
CAL	=	0.959259	0.800	<= CAL < 1.200
b1	=	0.0011	0.0000	<= b1 < 0.0040
b2	=	0.0055	0.0010	<= b2 < 0.0100
b3	=	0.0003	0.0000	<= b3 < 0.0040
Xg	=	0.1030	0.0500	<= Xg < 0.2000
a21	=	1.167183	1.050	<= a21 < 1.300
a31	=	0.555798	0.300	<= a31 < 0.800

*COB/2*

Final page for report

\

\\ANCSCLSAN01\JTRAX\CASEDOCS\B-100412\B-100412\_ECF-20140

2/18/2014