

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100366

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date:01/06/2022

## External Standard Test Values

## Diagnostic Check

### VERSIONS

DMT: 3.02

PIC: 3.03

Modem: 2.6

Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C

PASSED

Breath Tube = 44.2°C

PASSED

### PUMP INFO

Flow Rate = 4.768 L/M

PASSED

### DETECTOR INFO

PUMP ON

PASSED

PUMP OFF

FAILED

### FILTER INFO

Filter 1

PASSED

Filter 2

PASSED

Filter 3

PASSED

### INTERNAL STANDARD

FAILED

Diagnostic Check Failed