

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Issued: 04/03/2018
Effective: 04/04/2018

Version: DMEF2018 R0
Status: Active

Instrument # 100376

Date: 4/25/18

Analyst: Colleen OByant

1. Reason for instrument evaluation

"Hard disk error"

2. Visual Inspection

Clean

Replace missing screws

Cermetex Box is secure

Ensure modem cable present

Gas Valve upright or secured to wall

All tubing and connections are secure

3. O-ring inspection and ensure 2 O-rings in replacement bag

4. Troubleshooting/Repair

Uploaded via modem line without issue.
Cleared memory of evaluation VOC, ran
a supervisor test + successfully uploaded
it to a USB drive.

5. VOC (attached)

6. Internal Standard value $\leq 3.00\%$ Yes No

7. Verify software version

Return to Factory for Repair

Ready for Calibration

Ready for Certification

Proceed to Close Out

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100376

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 04/25/2018

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.70: 0.079
LOT #: AG634001
EXPIRATION: 12/05/2018
TANK PRESSURE: 430 psi

BLANK TEST	0.000	13:02
INTERNAL STANDARD	VERIFIED	13:02
Xq = 0.076	(1.90%)	
EXTERNAL STANDARD	0.077	13:03
X[1] = 0.0774 (-0.0003)	(0.0000)	
BLANK TEST	0.000	13:03
EXTERNAL STANDARD	0.078	13:04
X[1] = 0.0776 (-0.0001)	(-0.0004)	
BLANK TEST	0.000	13:05
EXTERNAL STANDARD	0.078	13:05
X[1] = 0.0776 (-0.0012)	(-0.0008)	
BLANK TEST	0.000	13:06
EXTERNAL STANDARD	0.078	13:06
X[1] = 0.0783 (-0.0004)	(-0.0006)	
BLANK TEST	0.000	13:07
EXTERNAL STANDARD	0.078	13:07
X[1] = 0.0782 (-0.0002)	(-0.0007)	
BLANK TEST	0.000	13:08

Average = 0.0778
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C	PASSED
(44.0°C - 52.0°C)	
Breath Tube = 47.1°C	PASSED
(38.0°C - 50.0°C)	

SETTINGS

Lamp Voltage = 1.59 V	PASSED
(0.98 V - 2.30 V)	
Cooler Voltage = 1.53 V	PASSED
(0.98 V - 2.28 V)	
Bias Voltage = 79 V	PASSED
(47 V - 111 V)	
Chopper Freq = 527 Hz	PASSED
(475 Hz - 575 Hz)	
Barometer = 29.7 in	

PUMP INFO

Flow Rate = 4.375 L/M	PASSED
(3.500 L/M - 6.500 L/M)	

DETECTOR INFO

PUMP ON	PASSED
(0.002 V <= 0.010 V)	
PUMP OFF	PASSED
(0.002 V <= 0.010 V)	

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
Xq = 0.076	(2.02%) (0.00% - 4.00%)