

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 05/01/2015 Instrument Serial Number: 100348

Instrument Location: SAND POINT P.D.

Supervisor Name and Agency: MICHAEL CHIESA, SAND POINT P.D.

Describe the Status Change or Issue with the Instrument:

<u>DETECTOR</u>	<u>OVERFLOW</u>	<u>STATUS</u>	<u>MESSAGE</u>
<u>INSTRUMENT</u>	<u>WILL</u>	<u>NOT</u>	<u>PASS</u>
			<u>V.I.O.C.</u>

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 05/01/2015 / 1550 hrs

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740