

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 12/15/2017  
Effective: 12/15/2017

Version: CIISF 2017 R0  
Status: Active

Date: 4/12/2018 Instrument Serial Number: 100690

Instrument Location: UNALASKA DEPT. OF PUBLIC SAFETY.

Supervisor Name and Agency: KEVIN WOOD UNALASKA

Describe the Status Change or Issue with the Instrument:

Replace DataMASTER 100376 which had a Hard Disk  
ERROR WHICH WAS UNABLE TO BE CLEARED.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 4/12/2018 13:03

Out of Service Date/Time: \_\_\_\_\_

For Use by SCDL

Additional Notes

---

---

---

---

---

---

---

---

---

---

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740