

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100666

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 04/27/2020

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.59: 0.079  
LOT #: AG826101  
EXPIRATION: 09/18/2020  
TANK PRESSURE: 609 psi

BLANK TEST	0.000	13:47
INTERNAL STANDARD	VERIFIED	13:47
EXTERNAL STANDARD	0.076	13:47
BLANK TEST	0.000	13:48
EXTERNAL STANDARD	0.076	13:48
BLANK TEST	0.000	13:49
EXTERNAL STANDARD	0.076	13:49
BLANK TEST	0.000	13:50
EXTERNAL STANDARD	0.077	13:50
BLANK TEST	0.000	13:51
EXTERNAL STANDARD	0.077	13:51
BLANK TEST	0.000	13:52

Average = 0.0764  
Std Dev = 0.0005

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.6°C PASSED  
Breath Tube = 41.3°C PASSED

### PUMP INFO

Flow Rate = 4.071 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED

### FILTER INFO

Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

INTERNAL STANDARD PASSED

*DMV*  
*File only*  
*DMV 01/31/24*