Alaska Scientific Crime Detection Laboratory Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 5/21/2025

Instrument #: 100695

Instrument Location: Sand Point PD / SCDL

Supervisor Name and Agency: Carlie Glaister - SCDL

Describe the Status Change or Issue with the Instrument:

This instrument was returned from Sand Point PD due to the crime lab not receiving datafiles from the agency while it was in service. Sand Point PD does not have a functioning modem line, and the agency was not manually sending the files to the crime lab. When the instrument was checked in there was 1 datafile that uploaded which seemed unusual. The file was a VOC for 04/11/2025. After inspecting the DataMaster reports section it was determined that there were more datafiles that had been removed from the short term memory, but were not received previously by the crime lab. In DMHost the Adjustment, Breath, Supervisor, Disgnostic, Tank Change, and VOC records were queried for all tests with a search start date of 11/20/2024. Only 1 record was present, the VOC from 4/11/2025 that was received during the check in upload. The SD card was removed from the instrument and used to retrieve all datafiles from the instrument. After retrieval DMHost was queried again for Adjustment, Breath, Supervisor, Diagnostic, Tank Change and VOC records using a start date of 11/20/2024. There were an additional 6 records that were retrieved, which were all VOCs: 2 VOCs from 11/23/2024, and 11/26/2024, 12/12/2024, 1/3/2025, and 2/21/2025. There were no records present for Adjustment, Breath, Supervisor, Diagnostic or Tank Change.

* - Correct spelling should be Diagnostic CKG 5/22/2025

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA

Out of Service Date/Time: NA

For Use by SCDL

Additional Notes

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

Page 1 of 2

Version: 6.0

Click or tap here to enter text.

Effective: 1/17/2024

Email completed form to dps.scdl.toxicology@alaska.gov.

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

Page 2 of 2