

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 03/19/2018 Instrument Serial Number: 100360

Instrument Location: AST/Soldotna

Supervisor Name and Agency: Ronny Simmons (via email)

Describe the Status Change or Issue with the Instrument:

Ronny performed a VOC on 100360 on 03/19/2018 at 10:52AM with the incorrect lot number and expiration date entered.
The lot number and expiration date have been corrected on the VOC.

Incorrect lot number and expiration date: AG634001 EXP 12/05/2018
Correct lot number and expiration date: AG608504 EXP 03/25/2018
(COB 03/19/2018)

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA

Out of Service Date/Time: NA

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

O'Bryant, Colleen S (DPS)

From: Simmons, Ronny A (DPS)
Sent: Monday, March 19, 2018 11:10 AM
To: DPS SCDL Toxicology (DPS sponsored)
Subject: tank change

To whom it may concern, I was doing a tank change at Soldotna AST due to an expiring bottle. I closed out of the screen where it said confirm tank is inserted to do something else thinking it would stop the process. Either way it started the VOC on the old bottle that was still in there but used the information I had typed in for the new bottle that hadn't been inserted. I have since put in the new bottle and completed the tank change.

The VOC that has the wrong bottle is from 3-19-18 at 1052 hours. The old bottle lot number is AG608504.

If this doesn't make sense give me a call. Take Care, Ronny