

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100342

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date:08/15/2021

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.85: 0.080  
LOT #: AG018102  
EXPIRATION: 06/29/2022  
TANK PRESSURE: 517 psi

BLANK TEST                      0.000    12:25  
INTERNAL STANDARD            VERIFIED 12:26

Gas Flow Error

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.7°C            PASSED  
Breath Tube       = 44.9°C           PASSED

### PUMP INFO

Flow Rate = 4.253 L/M               PASSED

### DETECTOR INFO

PUMP ON                                PASSED  
PUMP OFF                               PASSED

### FILTER INFO

Filter 1                                 PASSED  
Filter 2                                 PASSED  
Filter 3                                 PASSED

INTERNAL STANDARD                   PASSED