

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*  
State of Alaska

Serial #: 100406

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 01/03/2024

## External Standard Test Values

**EXTERNAL STANDARD INFORMATION**

NOMINAL: 0.080  
 TARGET AT 27.19: 0.073  
 LOT #: AG310901  
 EXPIRATION: 04/19/2025  
 TANK PRESSURE: 1138 psi

BLANK TEST	0.000	12:02
INTERNAL STANDARD	VERIFIED	12:02
EXTERNAL STANDARD	0.073	12:02
BLANK TEST	0.000	12:03
EXTERNAL STANDARD	0.073	12:03
BLANK TEST	0.000	12:04
EXTERNAL STANDARD	0.072	12:04
BLANK TEST	0.000	12:05
EXTERNAL STANDARD	0.072	12:05
BLANK TEST	0.000	12:06
EXTERNAL STANDARD	0.073	12:06
BLANK TEST	0.000	12:07

Average = 0.0726  
 Std Dev = 0.0005

## Diagnostic Check

**VERSIONS**  
 DMT: 3.02  
 PIC: 3.02  
 Modem: 2.6  
 Questions: 2.2

**TEMPERATURES**

Sample Chamber = 48.8°C	PASSED
Breath Tube = 40.9°C	PASSED

**PUMP INFO**  
 Flow Rate = 4.470 L/M PASSED


**DETECTOR INFO**  
 PUMP ON PASSED  
 PUMP OFF PASSED

**FILTER INFO**  
 Filter 1 PASSED  
 Filter 2 PASSED  
 Filter 3 PASSED


INTERNAL STANDARD PASSED


I, Derek J. Walton, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) As of the date signed below, I certify that the calibration of the referenced instrument was accurate on the date in which the above tests were performed and therefore certified for evidentiary use in the State of Alaska.

  
 Derek J. Walton  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 31 day of 01, 2024

  
 Notary Public  
 My Commission Expires With Office



0106  
 1/23/24