

**In Field Instrument Review**  
**Date Range: 2021.12.13 – 2022.01.03**  
**Review Completed on 2022.01.31 by Colleen O’Bryant**

**Status Messages and Action Taken**

**100355 (Homer PD)** – 1/8 RFI Status Message on Subject Tests – No action taken.

**100366 (Anchorage PD)** – 1/48 Blank Error Status Message on Subject Tests – Removed from service on 2022.01.07 (100366 2022.01.07 Change in Status) for internal standard status message on 2022.01.05 VOC.

**100378 (Palmer PD)** – 3/17 - RFI Status Message on Subject Tests – No action taken.

**100381 (AST/Talkeetna)** - 1/4 Tank Interference Status Messages on VOC and 1/4 Diagnostic Failed Status Messages on VOC due to Internal Standard Failure – Both status messages were followed by a successful test – Recommend monitoring instrument.

**100385 (AST/Fairbanks)** – 1/18 Tank Interference Status Message on Subject Tests – No action taken.

**100396 (Glacier Bay NP)** – 2/2 Diagnostic Failed Status Messages on VOCs due to Detector Overflow – Recommend Instrument be removed from service (100396 2022.01.06 Change in Status).

**100398 (Anchorage PD)** – 2/4 Blank Error Status Messages on VOCs – Tank Interference followed by blank error on back-to-back VOCs, followed by two successfully completed VOCs (2021.12.29, 2022.01.03) – Recommend Monitoring Instrument

**100402 (AST/Tok)** - 1/4 Tank Interference Status Message on Subject Tests – Followed by successful test and hasn’t reoccurred – No action taken.

**100421 (JBER)** – 3/17 RFI Status Message on Subject Tests, 1/17 Interference Status Messages on Subject Tests, 1/17 Blank Error Status Message on Subject Test – The instrument was not uploading properly so it was brought to SCDL for troubleshooting and returned to service. (100421 2022.01.10 Change In Status), (100421 2022.01.12 Change In Status).- No further action taken.

**100681 (Soldotna PD)** – 1/8 RFI Status Message on Subject Tests – No action taken.

**Follow Up from Previous In Field Instrument Review**

**100360 (Sand Point PD)** – 1/3 Internal Standard Error Status Message on Subject Test – Occurred on a “test/test” scenario. 1/2 Diagnostic Failed on VOC due to internal standard failure – followed by a successfully completed VOC (100360 2021.10.30 Change In Status) – the instrument was not removed from service. – Recommend Monitoring for 2021-10.25 – 2021.12.12 – Instrument removed from service by 2021.11.19. Instrument and CIISF are in transit as of 2022.01.16.

**Agencies not included in this summary:**

**100384 (Sand Point PD)** - Data is incomplete, last upload 2021.11.04

**100424 (Eielson AFB)** – No data received.

**100682 (Kodiak PD)** – Data is incomplete, last upload 2021.11.24

**100692 (Girdwood)** – No data received.

**100696 (Kake PD)** – Data is incomplete, last upload 2021.12.12

**Subject Test Data**

Serial Number	Total Subject Tests	RFI		Pump Error		Interference		Tank Interference		Date Range		Total Tests		Total Messages		Ambient Fail		Internal Standard Error		Diagnostic Failed					
										12/13/2021		1/3/2022		336		13		3.9%							
										Blank Error	Standard Out of Range	Gas Flow Error													
100355	8	1	13%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%				
100366	48	0	0%	0	0%	0	0%	0	0%	1	2%	0	0%	0	0%	0	0%	0	0%	0	0%				
100378	17	3	18%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%				
100385	18	0	0%	0	0%	0	0%	1	6%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%				
100402	4	0	0%	0	0%	0	0%	1	25%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%				
100421	17	3	18%	0	0%	1	6%	0	0%	1	6%	0	0%	0	0%	0	0%	0	0%	0	0%				
100681	8	1	13%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%				
<b>Total</b>	<b>336</b>	<b>8</b>	<b>2%</b>	<b>0</b>	<b>0%</b>	<b>1</b>	<b>0%</b>	<b>2</b>	<b>1%</b>	<b>2</b>	<b>1%</b>	<b>0</b>	<b>0%</b>	<b>0</b>	<b>0%</b>	<b>0</b>	<b>0%</b>	<b>0</b>	<b>0%</b>	<b>0</b>	<b>0%</b>				



VOC Data

Serial Number	Total VOC Tests	RFI		Pump Error				Interference				Tank Interference		Date Range		Total Tests				Total Messages		Diagnostic Failed			
														12/13/2021	1/3/2022	81		6		7.4%					
														Blank Error		Standard Out of Range		Gas Flow Error		Ambient Fail				Internal Standard Error	
100381	4	0	0%	0	0%	0	0%	0	0%	1	25%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	1	25%
100396	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	2	100%
100398	4	0	0%	0	0%	0	0%	0	0%	0	0%	2	50%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%
Total	81	0	0%	0	0%	0	0%	0	0%	1	1%	2	2%	0	0%	0	0%	0	0%	0	0%	0	0%	3	4%

