

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100418

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 02/24/2017

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 30.05: 0.080
LOT #: AG608504
EXPIRATION: 03/25/2018
TANK PRESSURE: 1307 psi

BLANK TEST	0.000	10:20
INTERNAL STANDARD	VERIFIED	10:20
EXTERNAL STANDARD	0.085	10:20
BLANK TEST	0.000	10:21
EXTERNAL STANDARD	0.085	10:22
BLANK TEST	0.000	10:23
EXTERNAL STANDARD	0.085	10:23
BLANK TEST	0.000	10:24
EXTERNAL STANDARD	0.085	10:25
BLANK TEST	0.000	10:26
EXTERNAL STANDARD	0.085	10:26
BLANK TEST	0.000	10:27

Average = 0.0850
Std Dev = 0.0000

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 50.5°C	PASSED
Breath Tube = 48.1°C	PASSED

PUMP INFO

Flow Rate = 4.533 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

File Only: Instrument removed
from service. CNF 3/7/17