

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100342

Received From: Bethel PD

Date: 5/28/24

Analyst: Kaija Swisher

### Reason for instrument return:

9/21 Pump error on subject tests (all on same subject), 3/21  
Subject interference on subject tests (all same subject), 3/21  
Internal standard on subject tests - replacement Rec. - Infield inst. Rev.

Change instrument and case location in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax

Upload

Update agency information and polling status in DMHost Calling Program

Check for VOCs

### From Intoximeters

Intoximeters Repair Paperwork (attached)