

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Issued: 04/03/2018
Effective: 04/04/2018

Version: DMEF2018 R0
Status: Active

Instrument # 100413

Date: 12/16/19

Analyst: Derek Walton

1. Reason for instrument evaluation

Sticky 5 way valve

2. Visual Inspection

- | | |
|--------------------------------------------------------------------------|---------------------------------------------------------------------------|
| <input checked="" type="checkbox"/> Clean | <input checked="" type="checkbox"/> Replace missing screws |
| <input checked="" type="checkbox"/> Cermetex Box is secure | <input checked="" type="checkbox"/> Ensure modem cable present |
| <input checked="" type="checkbox"/> Gas Valve upright or secured to wall | <input checked="" type="checkbox"/> All tubing and connections are secure |

3. O-ring inspection and ensure 2 O-rings in replacement bag

4. Troubleshooting/Repair

Replaced o-ring.

Routed filter wires separately.

Adjusted detector voltages.

5 way valve visibly sticking - replaced.

5. VOC (attached)

6. Internal Standard value $\leq 3.00\%$ Yes No

7. Verify software version

Return to Factory for Repair

Ready for Calibration

Ready for Certification

Proceed to Close Out

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100413

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/16/2019

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.20: 0.078
LOT #: AG923401
EXPIRATION: 08/22/2021
TANK PRESSURE: 540 psi

BLANK TEST	0.000	10:32
INTERNAL STANDARD ✓	VERIFIED	10:32
Xq = 0.094 (0.83%)		
EXTERNAL STANDARD	0.076	10:32
X[1] = 0.0756 (-0.0013) (0.0032)		
BLANK TEST	0.000	10:33
EXTERNAL STANDARD	0.075	10:33
X[1] = 0.0752 (-0.0028) (0.0028)		
BLANK TEST	0.000	10:34
EXTERNAL STANDARD	0.076	10:34
X[1] = 0.0755 (-0.0030) (0.0020)		
BLANK TEST	0.000	10:35
EXTERNAL STANDARD	0.076	10:35
X[1] = 0.0763 (-0.0031) (0.0023)		
BLANK TEST	0.000	10:36
EXTERNAL STANDARD	0.076	10:36
X[1] = 0.0760 (-0.0023) (0.0020)		
BLANK TEST	0.000	10:37

Average = 0.0758
Std Dev = 0.0004

Diagnostic Check

VERSIONS ✓
DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C	PASSED
(44.0°C - 52.0°C)	
Breath Tube = 43.8°C	PASSED
(38.0°C - 50.0°C)	

SETTINGS

Lamp Voltage = 1.98 V	PASSED
(1.08 V - 2.52 V)	
Cooler Voltage = 1.84 V	PASSED
(0.95 V - 2.23 V)	
Bias Voltage = 79 V	PASSED
(47 V - 111 V)	
Chopper Freq = 526 Hz	PASSED
(475 Hz - 575 Hz)	
Barometer = 29.2 in	

PUMP INFO

Flow Rate = 5.262 L/M	PASSED
(3.500 L/M - 6.500 L/M)	

DETECTOR INFO

PUMP ON	PASSED
(0.004 V <= 0.010 V)	
PUMP OFF	PASSED
(0.003 V <= 0.010 V)	

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD ✓	PASSED
Xq = 0.093 2.03% (0.00% - 4.00%)	