

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 12-28-2023

Instrument #: 100682

Instrument Location: King Cove Police Department

Supervisor Name and Agency: Chris Babcock King Cove DPS

Describe the Status Change or Issue with the Instrument:

Put this Datamaster into service

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: ~~12-28-2023 1500~~ 1/3/2024 (CKG 3/26/2024)

Out of Service Date/Time:

For Use by SCDL

Additional Notes

This agency is unable to poll. It was later discovered that the VOC on 12/28/2023 did not complete successfully due to a Diagnostic Failed associated with the pump. The first completed VOC occurred on 1/3/2024. This record is being amended to reflect the in service date of 1/3/2024. CKG 3/26/2024

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

100682 2024.12.28 Change In Status - Amended

100682 2023.12.28 Change In Status