

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100424

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/04/2023

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 28.66: 0.077
LOT #: AG310901
EXPIRATION: 04/19/2025
TANK PRESSURE: 1052 psi

BLANK TEST	0.000	04:53
INTERNAL STANDARD	VERIFIED	04:53
EXTERNAL STANDARD	0.079	04:53
BLANK TEST	0.000	04:54
EXTERNAL STANDARD	0.079	04:54
BLANK TEST	0.000	04:55
EXTERNAL STANDARD	0.079	04:55
BLANK TEST	0.000	04:56
EXTERNAL STANDARD	0.080	04:56
BLANK TEST	0.000	04:57
EXTERNAL STANDARD	INTERFERENCE	04:58

Interference Detected

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 46.7°C	PASSED

PUMP INFO

Flow Rate = 3.870 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

PASSED

File by Jm 03/26/24