

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 9-2-2020 Instrument Serial Number: 100690
Instrument Location: in Route To AK Crime Lab
Supervisor Name and Agency: Josh Lobato Bristol Bay Police Dept.

Describe the Status Change or Issue with the Instrument:

Sendin this unit to Alaska Crime Lab For
Routine Swap and Service

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: ~~9-2-20~~ 12:37 P.M SML

Out of Service Date/Time: 9-2-20 11:35 AM

For Use by SCDL

Additional Notes

DM 09/09/20
Instrument experiencing ~~blank~~ blank errors and internal standard
errors DM 09/09/20

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740