

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Serial #: 100360

Date: 04/13/2021

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.76: 0.080
LOT #: AG923401
EXPIRATION: 08/22/2021
TANK PRESSURE: 1044 psi

BLANK TEST	0.000	12:30
INTERNAL STANDARD	VERIFIED	12:30
EXTERNAL STANDARD	0.077	12:30
BLANK TEST	0.000	12:31
EXTERNAL STANDARD	0.078	12:31
BLANK TEST	0.000	12:32
EXTERNAL STANDARD	0.076	12:33
BLANK TEST	0.000	12:33
EXTERNAL STANDARD	0.078	12:34
BLANK TEST	0.000	12:35
EXTERNAL STANDARD	0.078	12:35
BLANK TEST	0.000	12:36

Average = 0.0774
Std Dev = 0.0009

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
Breath Tube = 44.9°C PASSED

PUMP INFO

Flow Rate = 4.720 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO


Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD

PASSED


I, Brandi M. Barnett, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.



 5/21/21

Brandi M. Barnett
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 21 day of 5, 2021



Notary Public
My Commission Expires With Office



JW 04/14/21