

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 10 May 2016 Instrument Serial Number: 100687
Instrument Location: JNEN JBR4
Supervisor Name and Agency: Scott Gardner - JBR4

Describe the Status Change or Issue with the Instrument:

Removed from SVS - Loner
- multiple interference messages
- won't pass VOC

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 4/19/2016

For Use by SCDL

Additional Notes

Email completed form to _____

For questions contact the Breath Alcohol Program at 907-269-5740

O'Bryant, Colleen S (DPS)

100687

From: GARDNER, SCOTT E GS-11 USAF PACAF 673 SFS/S3T <scott.gardner.14@us.af.mil>
Sent: Monday, April 18, 2016 3:26 PM
To: O'Bryant, Colleen S (DPS)
Subject: RE: 100687

Collen,

Sounds good. Let me know when and I'll swap them out.

Scott

-----Original Message-----

From: O'Bryant, Colleen S (DPS) [<mailto:colleen.obryant@alaska.gov>]
Sent: Monday, April 18, 2016 2:58 PM
To: GARDNER, SCOTT E GS-11 USAF PACAF 673 SFS/S3T <scott.gardner.14@us.af.mil>
Subject: RE: 100687

Scott,

i have your instrument (100421) warming up to certify this week. How about I give you a shout when it's done and we can trade?

Colleen

From: GARDNER, SCOTT E GS-11 USAF PACAF 673 SFS/S3T [scott.gardner.14@us.af.mil]
Sent: Monday, April 18, 2016 1:55 PM
To: O'Bryant, Colleen S (DPS)
Subject: RE: 100687

Colleen,

Did another VOC, and another, and another..... Looks like your suspicions are validated. It got several inference detected messages. I'm going to swap it out with our other Data Master 100422. Was planning on moving from our other building anyway so this is a good opportunity. Let me know when you'd like your loaner back.

Scott

Ofc Scott Gardner, GS-11, DAFCP
Training Officer
JBER Police
Joint Base Elmendorf-Richardson, AK
Work: 552-3537
Cell: 350-6140

-----Original Message-----

From: O'Bryant, Colleen S (DPS) [<mailto:colleen.obryant@alaska.gov>]

Sent: Thursday, April 14, 2016 3:36 PM

To: GARDNER, SCOTT E GS-11 USAF PACAF 673 SFS/S3T <scott.gardner.14@us.af.mil>

Subject: 100687

Scott,

Could you please perform a VOC on your loaner instrument (100687), the one performed on 04/11/2016 for VOC day completed, but I'd like to see results of another one.

Thanks,

Colleen O'Bryant - Forensic Scientist III

Blood and Breath Alcohol

Alaska Scientific Crime Detection Laboratory

4805 Dr. Martin Luther King Jr. Ave

Anchorage, AK 99507

907-269-5783 (direct)

907-350-2111 (cell)